

Microwave Spectrometer HK9

- Idea behind the spectrometer concept; purpose and application
- Physical background
- Realisation (instrument, antennas, preliminary technical data)
- Specter 9 software for calibration
- Indoor tests on different products (tobacco, sand, Isopropanol2, sugar)
- Task list; field test on tobacco, temperature tests, etc.



Idea behind the spectrometer concept

If it is required to measure materials that are baled or packed in boxes, the moisture and the density inside the bale or box are of interest.

Problems using phaseshift and attenuation of the μW signal:

- Phaseshift is only defined in a range of 0° to 360°
- Attenuation has a small measuring effect
- To detect density and moisture a combination of attenuation and phaseshift is needed

Idea behind the spectrometer concept

Advantage using the absorption spectrum:

- no limitation in measuring range (0° to 360° phaseshift)
- As the interaction of material and water is frequency dependent, moisture and density can be measured

Remark: As the attenuation shows less measuring effect than the phase shift, the spectrometer is not usable for small quantities of materials with low moisture.

- Use frequency dependent attenuation instead of phase measurement

Physical background

The interaction of material and microwaves is specified by the dielectric constant of the material, epsilon.

$$\epsilon = \epsilon(\text{real}) + \epsilon(\text{imaginary})$$

$$\epsilon(\text{real}) = \text{reduction in wavelength}$$

$$\epsilon(\text{imaginary}) = \text{angle of energy absorption}$$

The loss tangent

$$\tan(\delta) = \epsilon(\text{imaginary}) / \epsilon(\text{real})$$

describes the loss in dielectrics and is the attenuation measured by the HK9 instrument. The variation in $\tan(\delta)$ relates to the moisture content and the bulk density of the material.

Physical background

The dielectric behaviour of material, especially of water, is frequency dependent . This factor enables us to generate a density-independent moisture signal.

Using a multilinear regression of the absorption spectrum makes it possible to distinguish between the attenuation caused by moisture and that caused by density.

The standardisation, used in the HK9 software, also helps generating a density-independent signal. The first derivation, dT/λ , of the spectrum is one of the possibilities the HK9 software offers.

HK9 Technical data (priliminary)

Housing: Aluminium; IP65; H*B*T = 230*200*110mm

Weight: about 5kg

Voltage: 115/230 VAC +/-15% 47-65 Hz

Power consumption: 50VA

Operating temperature: -20 – 85 °C

Storage temperature:-30 – 95 °C

Display: 2*24 character LCD, LED-backlight

Relay: AC 250VA, DC 30V 1A

Current output: 2 * 0/4 – 20mA active output, isolated,
max. load 500 Ohm, moisture and
density

HK9 Technical data (priliminary)

Current input: 0/4 – 20mA, not isolated, load 125 Ohm

COM ports: RS232; data output

RS232; bidirectional; calibration via SPECTER9

RS485; data output

2400 Baud to 115 kBaud;


8 data bit, 1 stop bit, no parity

Digital inputs: Start/Stop; Product select

Microwave frequency: 1,5 GHz to 3,5 GHz; 50 steps

Microwave power: max: -10 dBm

Sensitivity: -80 dBm; 10 measurements/second



Connection Diagram HK9



**HARRER
KASSEN**


Harrer & Kassen GmbH
An Hieschen 6
D-75328 Langenbrand

Typ : HK9
Tobacco


AC : 100-240V 50/60 Hz
200 mA max.

Made in Germany
www.harrer-kassen.com Tel: +49 (0)7064-5245-0






Output power




Output power



Vivaldi group antenna

Dimension: 160*120*140 mm; L*B*H

Material: ABS; IP 65




Vivaldi antenna


Gain: 10 dB

HPBW: 60° to 70°

Radiation pattern Azimuth group



Radiation pattern Azimuth group



Specter 9 software for calibration

- Written by Mr. Gerald Keck
- Based on Spekter 8 Software for NIR calibration
- Uses multilinear Regression for calibrating HK9
- Calculates the coefficients for moisture and density
- PC based software that communicates via RS232 with the HK9 instrument
- Calculates up to 6 coefficients for moisture and up to 7 coefficients for density measurement.

Specter 9 software for calibration

$$H_2O = a_0 + a_1 * W_1 + a_2 * W_2 + a_3 * W_3 + a_4 * W_4$$

with W (1->n) standardised density independent values

$$Rho = a_0 + a_1 * T_1 + a_2 * T_2 + a_3 * T_3 + a_4 * T_4 \\ + a_5 * T_5 + a_6 * H_2O$$

with T (1->n) standardised spectrum

Remark: Number of needed coefficients reduces, if less than 5 wavelength are used.



Status

COM-Port: OK. | Calculation: OK.



Calibration

B0: 5.45 B1: -9.891759 B2: -13.30408 B3: -3.036403 B4: -0.269627 B5: 0.000000

Product: Name: Standardisation: DLG

Calibration quality

STD: Calibration: 0.070 STD: Unknown samples: -NA- $R^2: 0.800$


Reference

W1: W2: W3: W4: W5:

Sample View


x	Comment	W1	W2	W3	W4	W5	H2O	P.H2O	E.H2O
<input checked="" type="checkbox"/>	09.04.13 13:54	54.22	48.03	32.38	35.13	36.31	1.00	0.87	-0.13
<input checked="" type="checkbox"/>	09.04.13 13:59	54.22	48.03	32.28	35.59	35.77	1.00	0.87	-0.13
<input checked="" type="checkbox"/>	09.04.13 14:12	54.47	48.04	32.54	35.42	36.29	0.95	0.91	-0.04
<input checked="" type="checkbox"/>	09.04.13 14:15	54.39	48.00	32.51	35.43	36.27	0.95	0.93	-0.02
<input checked="" type="checkbox"/>	09.04.13 14:22	54.79	48.25	32.68	35.38	36.39	0.90	0.84	-0.06
<input checked="" type="checkbox"/>	09.04.13 14:24	54.74	48.21	32.65	35.33	36.28	0.90	0.84	-0.06
<input checked="" type="checkbox"/>	09.04.13 14:27	55.03	48.47	32.74	35.60	36.47	0.85	0.79	-0.06
<input checked="" type="checkbox"/>	09.04.13 14:29	54.93	48.36	32.64	35.60	36.46	0.85	0.78	-0.07
<input checked="" type="checkbox"/>	09.04.13 13:54	56.64	50.18	34.41	36.45	37.35	1.00	0.97	-0.03
<input checked="" type="checkbox"/>	09.04.13 14:12	57.09	50.55	34.58	37.17	37.46	0.95	0.95	0.00
<input checked="" type="checkbox"/>	09.04.13 14:15	57.07	50.57	34.57	37.39	37.46	0.95	0.98	0.03
<input checked="" type="checkbox"/>	09.04.13 14:22	57.38	50.80	34.84	36.99	37.49	0.90	0.93	0.03
<input checked="" type="checkbox"/>	09.04.13 14:24	57.37	50.74	34.76	37.02	37.38	0.90	0.90	0.00
<input checked="" type="checkbox"/>	09.04.13 14:27	57.75	51.11	35.14	37.35	37.92	0.85	0.96	0.11
<input checked="" type="checkbox"/>	09.04.13 14:29	57.71	51.09	35.09	37.22	37.83	0.85	0.94	0.09

Lab versus HK9




Calibration quality:

Deviation



XY View



Count:

Specter 9 software for calibration

Calibration

B0: 5.45 B1: -9.891759 B2: -13.30408 B3: -3.036403 B4: -0.269627 B5: 0.000000

Product: H2O Name: Isopropanol2 Standardisation: dT/Lambda DLG

Send Calib

Coeffitions water

Calibration water

Standardisation

T ; dT/Lambda ;
Special ; Linear

Calibration

B0: 0.53 B1: -0.006161 B2: -0.002711 B3: 0.003081 B4: 0.027312 B5: 0.014034
B6: 0.100206

Product: Rho2 Name: Isopropanol2 Standardisation: T DLG

Send Calib

Coeffitions density

Calibration density

Calibration quality

STD: Calibration: 0.070 STD: Unknown samples: -NA- R^2 : 0.800

Statistical parameters; Quality of calibration

Specter 9 software for calibration

Reference

W1	W2	W3	W4	W5	Send Ref.
21.000000	17.000000	8.000000	13.000000	9.000000	Edit Ref. Read Ref.HK9



Reference Attenuation for single wavelenth

Spectra with Lab-values; predicted values and deviation

x	Comment	W1	W2	W3	W4	W5	H2O	P.H2O	E.H2O
<input checked="" type="checkbox"/>	09.04.13 13:54	54.22	48.03	32.38	35.13	36.31	1.00	0.87	-0.13
<input checked="" type="checkbox"/>	09.04.13 13:59	54.22	48.03	32.28	35.59	35.77	1.00	0.87	-0.13
<input checked="" type="checkbox"/>	09.04.13 14:12	54.47	48.04	32.54	35.42	36.29	0.95	0.91	-0.04
<input checked="" type="checkbox"/>	09.04.13 14:15	54.39	48.00	32.51	35.43	36.27	0.95	0.93	-0.02
<input checked="" type="checkbox"/>	09.04.13 14:22	54.79	48.25	32.68	35.38	36.39	0.90	0.84	-0.06
<input checked="" type="checkbox"/>	09.04.13 14:24	54.74	48.21	32.65	35.33	36.28	0.90	0.84	-0.06
<input checked="" type="checkbox"/>	09.04.13 14:27	55.03	48.47	32.74	35.60	36.47	0.85	0.79	-0.06
<input checked="" type="checkbox"/>	09.04.13 14:29	54.93	48.36	32.64	35.60	36.46	0.85	0.78	-0.07
<input checked="" type="checkbox"/>	09.04.13 13:54	56.64	50.18	34.41	36.45	37.35	1.00	0.97	-0.03
<input checked="" type="checkbox"/>	09.04.13 14:12	57.09	50.55	34.58	37.17	37.46	0.95	0.95	0.00
<input checked="" type="checkbox"/>	09.04.13 14:15	57.07	50.57	34.57	37.39	37.46	0.95	0.98	0.03
<input checked="" type="checkbox"/>	09.04.13 14:22	57.38	50.80	34.84	36.99	37.49	0.90	0.93	0.03
<input checked="" type="checkbox"/>	09.04.13 14:24	57.37	50.74	34.76	37.02	37.38	0.90	0.90	0.00
<input checked="" type="checkbox"/>	09.04.13 14:27	57.75	51.11	35.14	37.35	37.92	0.85	0.96	0.11
<input checked="" type="checkbox"/>	09.04.13 14:29	57.71	51.09	35.09	37.22	37.83	0.85	0.94	0.09



Graphs

Moisture Density
Raw values





Standardisation

dT/Λ T



Comparison Lab vs. HK9

Moisture Density







Graphs

Moisture

Density

Deviation



Regression

Indoor tests on different products


- Isopropanol₂
- Sugar
- Sand
- Tobacco

HK9 Isopropanol2 Measurement


2 PIN-antennas tunable in distance, 4cm to 8cm

Bucket with water diluted with Isopropanol2, 100% to 50%

Changing the distance of the antennas corresponds to density variation



Lab versus HK9




Isopropanol2

dT/Lambda

Correlation=80%

STDeviation=7%


XY View



Isopropanol2

density variation 100%


Lab versus HK9



Isopropanol2

T Correlation=99,1% STDeviation=0,012=0,48mm

XY View



Isopropanol2

moisture variation 50%

HK9 Sugar Measurement

Sugar in a basket measured with two different sensors, spacing of the antennas 50mm and 75mm (picture)

Density independent standardisation provides calibration



that fits for both sensors.

1. Calibration 50mm spacing
2. Calibration all values 50mm and 75mm spacing
3. Calibration extreme values high/low



Status

COM-Port: OK. | Calculation: OK.



Calibration

B0: -191.96 B1: 176.3240 B2: 31.04700 B3: 210.0511 B4: 91.41023 B5: 0.000000

Product: Name: Standardisation: DLG

Calibration quality

STD: Calibration: 0.500 STD: Unknown samples: 17.099 R²: 0.996

Reference


W1	W2	W3	W4	W5	<input type="button" value="Send Ref."/>
<input type="text" value="10.000000"/>	<input type="text" value="25.000000"/>	<input type="text" value="1.000000"/>	<input type="text" value="20.000000"/>	<input type="text" value="10.000000"/>	<input type="button" value="Edit Ref."/>
					<input type="button" value="Read Ref.HK9"/>

Sample View

x	Comment	W1	W2	W3	W4	W5	H2O	P.H2O	E.H2O
✓	12.04.13 14:50	59.52	52.76	37.53	39.02	39.51	10...	99.68	-0.32
✓	12.04.13 14:51	59.58	52.69	37.53	39.04	39.51	10...	100.44	0.44
✓	12.04.13 14:54	59.96	53.03	37.76	39.26	39.60	97.50	97.54	0.04
✓	12.04.13 14:55	59.97	53.06	37.77	39.27	39.62	97.50	97.18	-0.32
✓	12.04.13 14:58	60.30	53.34	37.95	39.45	39.71	95.00	94.86	-0.14
✓	12.04.13 15:00	60.36	53.33	37.98	39.48	39.72	95.00	95.10	0.10
✓	12.04.13 15:07	60.68	53.57	38.17	39.69	39.86	92.50	92.65	0.15
✓	12.04.13 15:07	60.67	53.58	38.17	39.71	39.84	92.50	92.19	-0.31
✓	12.04.13 15:10	60.98	53.85	38.37	39.91	39.99	90.00	89.69	-0.31
✓	12.04.13 15:11	61.01	53.87	38.37	39.91	39.99	90.00	89.68	-0.32
✓	12.04.13 15:13	61.32	54.14	38.56	40.10	40.11	87.50	87.36	-0.14
✓	12.04.13 15:13	61.31	54.07	38.54	40.08	40.10	87.50	88.15	0.65
✓	12.04.13 15:18	61.60	54.30	38.73	40.28	40.22	85.00	85.97	0.97
✓	12.04.13 15:18	61.62	54.32	38.73	40.28	40.24	85.00	85.86	0.86
✓	12.04.13 15:22	62.27	54.77	39.08	40.69	40.39	82.50	81.90	-0.60


Count:

Lab versus HK9




Calibration quality:

Deviation




XY View





Calibration for 50mm sensor; 0 to 27 Bx

Lab versus HK9



Calibration for 50mm sensor



Two groups are visible representing different sensor spacings.

The normalized values show two groups, too. The amplitude of the two groups changes from + to -. As the graphs cross each other it is possible to combine them into one calibration.



Status

COM-Port: OK. | Calculation: OK.



Calibration

B0: -215.33 B1: -69.37059 B2: 311.1459 B3: 1012.274 B4: 1367.540 B5: 0.000000

Product: Name: Standardisation: DLG

Calibration quality

STD: Calibration: 2.106 STD: Unknown samples: 1.619 $R^2: 0.935$

Reference


W1	W2	W3	W4	W5	
<input type="text" value="10.000000"/>	<input type="text" value="25.000000"/>	<input type="text" value="1.000000"/>	<input type="text" value="20.000000"/>	<input type="text" value="10.000000"/>	<input type="button" value="Send Ref."/>
					<input type="button" value="Edit Ref."/>
					<input type="button" value="Read Ref.HK9"/>

Sample View

x	Comment	W1	W2	W3	W4	W5	H2O	P.H2O	E.H2O
<input checked="" type="checkbox"/>	10.04.13 15:39	72.97	64.09	46.73	46.70	45.09	90.00	91.62	1.62
<input checked="" type="checkbox"/>	10.04.13 15:40	73.59	64.59	47.14	47.10	45.42	87.50	88.42	0.92
<input checked="" type="checkbox"/>	10.04.13 15:42	73.87	64.64	47.27	47.20	45.47	87.50	88.54	1.04
<input checked="" type="checkbox"/>	10.04.13 15:44	74.06	64.91	47.57	47.51	45.75	85.00	84.34	-0.66
<input checked="" type="checkbox"/>	10.04.13 15:45	74.60	65.17	47.65	47.60	45.80	85.00	84.71	-0.29
<input checked="" type="checkbox"/>	10.04.13 15:47	74.64	65.42	47.98	47.87	46.05	82.50	83.45	0.95
<input checked="" type="checkbox"/>	10.04.13 15:48	75.46	65.45	48.08	48.03	46.12	82.50	79.64	-2.86
<input checked="" type="checkbox"/>	10.04.13 15:51	76.02	65.81	48.50	48.37	46.39	80.00	79.18	-0.82
<input checked="" type="checkbox"/>	10.04.13 15:52	76.20	65.91	48.54	48.45	46.41	80.00	78.66	-1.34
<input checked="" type="checkbox"/>	10.04.13 15:54	76.91	66.33	48.95	48.74	46.63	77.50	80.82	3.32
<input type="checkbox"/>	letzt	77.19	66.33	49.08	48.92	46.76	77.00	76.57	-0.43
<input type="checkbox"/>	letzt	77.39	66.37	49.15	48.96	46.81	77.00	76.36	-0.64
<input type="checkbox"/>	10.04.13 16:06	77.11	66.26	49.20	49.00	46.86	77.00	74.99	-2.01
<input type="checkbox"/>	10.04.13 16:08	77.62	66.55	49.14	48.93	46.82	77.00	77.67	0.67
<input type="checkbox"/>	10.04.13 16:24	77.21	66.42	49.10	48.87	46.75	77.00	79.30	2.30


Count:

Lab versus HK9




Calibration quality:

Deviation



XY View



Calibration 50mm and 75mm; all samples



Status

COM-Port: OK. | Calculation: OK.



Calibration

B0: -219.71 B1: -58.19203 B2: 319.4107 B3: 1021.306 B4: 1400.463 B5: 0.000000

Product: Name: Standardisation: DLG

Calibration quality

STD: Calibration: 1.642 STD: Unknown samples: 2.270 $R^2: 0.974$


Reference

W1	W2	W3	W4	W5	<input type="button" value="Send Ref."/>
<input type="text" value="10.000000"/>	<input type="text" value="25.000000"/>	<input type="text" value="1.000000"/>	<input type="text" value="20.000000"/>	<input type="text" value="10.000000"/>	<input type="button" value="Edit Ref."/>
					<input type="button" value="Read Ref.HK9"/>

Sample View


x	Comment	W1	W2	W3	W4	W5	H2O	P.H2O	E.H2O
<input type="checkbox"/>	12.04.13 15:10	60.98	53.85	38.37	39.91	39.99	90.00	87.75	-2.25
<input type="checkbox"/>	12.04.13 15:11	61.01	53.87	38.37	39.91	39.99	90.00	87.93	-2.07
<input type="checkbox"/>	12.04.13 15:13	61.32	54.14	38.56	40.10	40.11	87.50	85.98	-1.52
<input type="checkbox"/>	12.04.13 15:13	61.31	54.07	38.54	40.08	40.10	87.50	85.55	-1.95
<input type="checkbox"/>	12.04.13 15:18	61.60	54.30	38.73	40.28	40.22	85.00	82.78	-2.22
<input type="checkbox"/>	12.04.13 15:18	61.62	54.32	38.73	40.28	40.24	85.00	82.36	-2.64
<input type="checkbox"/>	12.04.13 15:22	62.27	54.77	39.08	40.69	40.39	82.50	78.01	-4.49
<input type="checkbox"/>	12.04.13 15:22	62.25	54.72	39.08	40.69	40.39	82.50	77.46	-5.04
<input type="checkbox"/>	12.04.13 15:26	62.64	55.13	39.32	40.93	40.52	80.00	77.06	-2.94
<input checked="" type="checkbox"/>	12.04.13 15:26	62.59	55.08	39.32	40.93	40.51	80.00	76.89	-3.11
<input type="checkbox"/>	12.04.13 15:30	62.94	55.39	39.54	41.11	40.60	77.50	78.51	1.01
<input type="checkbox"/>	12.04.13 15:30	62.95	55.32	39.53	41.12	40.60	77.50	76.83	-0.67
<input type="checkbox"/>	12.04.13 15:36	63.36	55.72	39.75	41.32	40.73	75.00	77.33	2.33
<input type="checkbox"/>	12.04.13 15:36	63.27	55.60	39.75	41.34	40.73	75.00	75.29	0.29
<input checked="" type="checkbox"/>	12.04.13 15:39	63.29	55.67	39.78	41.35	40.72	75.00	77.20	2.20

Lab versus HK9




Calibration quality:

Deviation




XY View



Count:

Calibration 50mm and 75mm; extreme values


Lab versus HK9



Calibration 50mm and 75mm

- all values

Lab versus HK9



- extreme values

Remark: The multilinear regression does not predict values that are outside the range used for calibration!

HK9 measurement on sand





HK9 Sand

HK9 measurement on sand

Two Vivaldi antennas, distance 50cm

Plastic pan 56 * 36 * 14 cm around 14 l

H₂O range 0 to 9%

Density; the weight of the plastic pan filled up with the sand was taken as density value. The variation in weight was 28 kg to 37,7 kg sand. Only one sample had 20 kg.

The sand was splitted into three parts, water was added and the moisture analyzed. After that two parts were filled into the pan and measured. The third part was added in three portions to give the density variation.

The layer thickness of the sand changed from 10cm to 14 cm.



Status

COM-Port: OK. | Calculation: OK.



Calibration

B0: 5.15 B1: 5.104699 B2: -1.649854 B3: 13.98629 B4: 18.01647 B5: 0.000000

Product: H2O Name: sand vivaldi Standardisation: dT/Lambda DLG

Send Calib

Calibration quality

STD: Calibration: 0.614 STD: Unknown samples: -NA- R²: 0.936

Reference

W1 W2 W3 W4 W5 Send Ref.
 40.000000 10.000000 10.000000 10.000000 10.000000 Edit Ref. Read Ref. HK9

Sample View

x	Comment	W1	W2	W3	W4	W5	H2O	P.H2O	E.H2O
<input type="checkbox"/>	18.04.13 15:58	44.98	42.06	31.52	30.29	30.16		0.99	
<input type="checkbox"/>	18.04.13 15:59	45.55	41.55	30.75	31.69	30.77		0.28	
<input type="checkbox"/>	18.04.13 16:05	45.25	40.69	29.88	32.53	30.84		-0.16	
<input type="checkbox"/>	18.04.13 16:08	45.21	40.60	30.24	31.94	30.09		0.65	
<input checked="" type="checkbox"/>	18.04.13 16:12	45.37	40.65	30.96	31.31	29.59	1.38	1.53	0.15
<input checked="" type="checkbox"/>	18.04.13 16:17	45.78	41.34	30.73	30.26	28.59	1.38	2.09	0.71
<input checked="" type="checkbox"/>	19.04.13 08:43	44.83	41.92	31.21	30.14	30.11	0.00	0.75	0.75
<input checked="" type="checkbox"/>	19.04.13 08:57	52.09	46.20	34.40	32.22	28.55	4.62	5.89	1.27
<input checked="" type="checkbox"/>	19.04.13 08:58	52.33	45.05	33.15	32.25	28.47	4.62	5.25	0.63
<input checked="" type="checkbox"/>	19.04.13 09:03	50.57	43.48	33.04	30.25	28.16	4.62	4.91	0.29
<input checked="" type="checkbox"/>	19.04.13 09:05	50.02	43.04	32.90	31.03	29.05	4.62	3.98	-0.64
<input checked="" type="checkbox"/>	19.04.13 09:10	49.95	44.04	32.97	30.71	29.24	4.62	3.65	-0.97
<input checked="" type="checkbox"/>	19.04.13 09:12	50.38	44.75	33.76	31.13	29.39	4.62	4.16	-0.46
<input checked="" type="checkbox"/>	19.04.13 09:16	49.78	43.21	31.79	29.19	27.62	4.62	4.18	-0.44
<input checked="" type="checkbox"/>	19.04.13 09:17	49.66	43.21	31.92	29.37	27.74	4.62	4.18	-0.44

Delete Sample

Edit Sample


New Sample manual

New Sample HK9

Count: 1

Exit


Lab versus HK9




Change view

Calibration quality:


Deviation



XY View




HK9 Raw values




Standardisation:
 dT/Λ

Calibration: H₂O

HK9 Normalized values



Lab versus HK9




Calibration: H2O


Correlation: 93,6%

STDeviation: 0,61%

Deviation



XY View





Status

COM-Port: OK. | Calculation: OK.



Calibration

B0: 5.00 B1: 0.064408 B2: 0.061495 B3: -0.228291 B4: -0.173749 B5: 0.206677
 B6: 0.063844

Product: Rho2 Name: sand vivaldi Standardisation: T DLG

Send Calib

Calibration quality

STD: Calibration: 0.164 STD: Unknown samples: -NA- R²: 0.738

Reference

W1 W2 W3 W4 W5 Send Ref.


40.000000 10.000000 10.000000 10.000000 10.000000 Edit Ref. Read Ref. HK9

Sample View

x	Comment	W1	W2	W3	W4	W5	Rho	P.H2O	P.Rho	E.Rho
<input type="checkbox"/>	18.04.13 ...	44.98	42.06	31.52	30.29	30.16		0.99	3.08	
<input type="checkbox"/>	18.04.13 ...	45.55	41.55	30.75	31.69	30.77		0.28	3.10	
<input type="checkbox"/>	18.04.13 ...	45.25	40.69	29.88	32.53	30.84		-0.16	3.07	
<input type="checkbox"/>	18.04.13 ...	45.21	40.60	30.24	31.94	30.09		0.65	2.98	
<input checked="" type="checkbox"/>	18.04.13 ...	45.37	40.65	30.96	31.31	29.59	2.77	1.53	2.88	0.11
<input checked="" type="checkbox"/>	18.04.13 ...	45.78	41.34	30.73	30.26	28.59	2.99	2.09	2.98	-0.01
<input type="checkbox"/>	19.04.13 ...	44.83	41.92	31.21	30.14	30.11		0.75	3.09	
<input checked="" type="checkbox"/>	19.04.13 ...	52.09	46.20	34.40	32.22	28.55	2.69	5.89	2.70	0.01
<input checked="" type="checkbox"/>	19.04.13 ...	52.33	45.05	33.15	32.25	28.47	2.70	5.25	2.91	0.21
<input checked="" type="checkbox"/>	19.04.13 ...	50.57	43.48	33.04	30.25	28.16	2.90	4.91	3.01	0.11
<input checked="" type="checkbox"/>	19.04.13 ...	50.02	43.04	32.90	31.03	29.05	2.90	3.98	3.03	0.13
<input checked="" type="checkbox"/>	19.04.13 ...	49.95	44.04	32.97	30.71	29.24	3.11	3.65	3.16	0.05
<input checked="" type="checkbox"/>	19.04.13 ...	50.38	44.75	33.76	31.13	29.39	3.11	4.16	3.01	-0.10
<input checked="" type="checkbox"/>	19.04.13 ...	49.78	43.21	31.79	29.19	27.62	3.40	4.18	3.30	-0.10
<input checked="" type="checkbox"/>	19.04.13 ...	49.66	43.21	31.92	29.37	27.74	3.40	4.18	3.26	-0.14

Delete Sample Edit Sample New Sample manual New Sample HK9


Lab versus HK9




Change view

Calibration quality:

Deviation




XY View




Count: 1 Exit

Lab versus HK9




Standardisation: T
Calibration: Rho2
Correlation: 73,8%
STDeviation: 0,16 kg

Deviation



XY View



HK9 measurement on sand

Moisture measurements on sand with variation in density and layer thickness are possible.

The amount of sand between the antennas should be at least 15cm to achieve good resolution and accuracy.

It is possible to measure the amount of material positioned between the antennas. This relates to density and load.

HK9 measurements on tobacco





HK9 tobacco

HK9 measurements on tobacco

- Two Vivaldi antennas, distance 50cm
- Cardboard box: size 40 * 60 * 30 cm around 72 l (1000g)
- H₂O range 10 to 25%
- Two types of tobacco: tobacco leaves and cutted brown tobacco
- Cutted tobacco: weight 2700g; moisture 10%
- Tobacco leaves: weight 2500g; moisture 16,5%
- Density lab value is weight of tobacco with cardboard box

In order to generate different moisture and density values, the tobacco leaves were put into plastic bags that were added to the cutted tobacco. (side by side or laying upon each other)

Time for absorbing water: at least 2 hours

cutted tobacco (old)

tobacco leaves (new)





Status

COM-Port: OK. | Calculation: OK.



Calibration

B0: 20.47 B1: 23.58084 B2: -0.689891 B3: -6.066230 B4: -3.248852 B5: 0.000000

Product: H2O Name: offsetvivaldi Standardisation: dT/Lambda DLG

Send Calib

Calibration quality

STD: Calibration: 1.621 STD: Unknown samples: -NA- R²:0.883

Reference


W1 W2 W3 W4 W5 Send Ref. Edit Ref. Read Ref. HK9

42.140000 40.540000 25.590000 25.580000 28.250000

Sample View

x	Comment	W1	W2	W3	W4	W5	H2O	P.H2O	E.H2O
✓	karton	42.26	40.77	25.87	25.68	28.61	0.00	0.74	0.74
✓	karton	42.17	40.71	25.91	25.85	28.58	0.00	0.84	0.84
✓	alt	42.60	41.20	25.86	26.14	28.63	12.50	13.93	1.43
✓	alt	42.85	41.30	26.25	26.27	28.54	12.50	14.60	2.10
✓	alt	42.76	41.28	26.13	26.23	28.50	12.50	12.22	-0.28
✓	neu	42.95	41.43	26.29	26.35	28.82	14.00	17.57	3.57
✓	neu	42.95	41.43	26.29	26.35	28.82	14.00	17.57	3.57
✓	neu	42.74	41.28	26.03	26.20	28.68	14.00	15.86	1.86
✓	alt+neu	43.25	43.03	27.31	26.89	29.55	13.00	15.96	2.96
✓	alt+neu	43.90	42.80	27.23	26.95	29.17	13.00	12.21	-0.79
✓	alt+neu	44.02	42.90	27.19	26.88	29.87	13.00	14.59	1.59
✓	alt+neu	44.34	43.18	27.34	27.17	29.10	13.00	12.75	-0.25
✓	alt	43.56	42.23	26.56	26.89	29.27	16.00	16.84	0.84
✓	alt	43.69	42.28	26.74	26.89	29.05	16.00	16.21	0.21


Lab versus HK9




Change view

Calibration quality:

Deviation



XY View




Count: 1

Exit


old

new

Combination of cutted tobacco and tobacco leaves



HK9 Raw values




Standardisation: dT/Λ

Calibration: H₂O


23% H₂O; old+new;
high density

HK9 Normalized values



Cardboard box

Lab versus HK9




Calibration: H₂O


Correlation: 88,3%

STDeviation: 1,6%

Deviation



XY View





Status

COM-Port: OK. | Calculation: OK.



Calibration

B0: 0.13 B1: -0.204423 B2: 0.323003 B3: 0.319120 B4: -0.420258 B5: -0.096581
 B6: 0.018132

Product: **Rho2** Name: offsetvivaldi Standardisation: T DLG

Send Calib

Calibration quality

STD: Calibration: 0.059 STD: Unknown samples: -NA- R²:0.866

Change view


Reference

W1 W2 W3 W4 W5
 42.140000 40.540000 25.590000 25.580000 28.250000

Sample View


x	Comment	W1	W2	W3	W4	W5	Rho	P.H2O	P.Rho	E.Rho
<input checked="" type="checkbox"/>	karton	42.26	40.77	25.87	25.68	28.61	0.10	0.74	0.19	0.09
<input checked="" type="checkbox"/>	karton	42.17	40.71	25.91	25.85	28.58	0.10	0.84	0.14	0.04
<input checked="" type="checkbox"/>	alt	42.60	41.20	25.86	26.14	28.63	0.38	13.93	0.29	-0.09
<input checked="" type="checkbox"/>	alt	42.85	41.30	26.25	26.27	28.54	0.38	14.60	0.35	-0.03
<input checked="" type="checkbox"/>	alt	42.76	41.28	26.13	26.23	28.50	0.38	12.22	0.34	-0.04
<input checked="" type="checkbox"/>	neu	43.15	41.77	26.51	26.39	28.77	0.35	13.38	0.48	0.13
<input checked="" type="checkbox"/>	neu	42.95	41.43	26.29	26.35	28.82	0.35	17.57	0.35	0.00
<input checked="" type="checkbox"/>	neu	43.02	41.71	26.31	26.48	28.86	0.35	13.86	0.37	0.02
<input checked="" type="checkbox"/>	neu	42.74	41.28	26.03	26.20	28.68	0.35	15.86	0.34	-0.01
<input checked="" type="checkbox"/>	alt+neu	44.39	43.03	27.31	26.89	29.55	0.63	15.96	0.58	-0.05
<input checked="" type="checkbox"/>	alt+neu	43.90	42.80	27.23	26.95	29.17	0.63	12.21	0.59	-0.04
<input checked="" type="checkbox"/>	alt+neu	44.02	42.90	27.19	26.88	29.87	0.63	14.59	0.55	-0.08
<input checked="" type="checkbox"/>	alt+neu	44.34	43.18	27.34	27.17	29.10	0.63	12.75	0.58	-0.05
<input checked="" type="checkbox"/>	alt	43.56	42.23	26.56	26.89	29.27	0.40	16.84	0.34	-0.06
<input checked="" type="checkbox"/>	alt	43.69	42.28	26.74	26.89	29.05	0.40	16.21	0.40	0.00

Lab versus HK9




Calibration quality:

Deviation



XY View



HK9 measurements on tobacco

Measurement of H₂O and density is possible.

5 kg tobacco is too little amount to get better results in moisture and density.

The amount of water between the antennas correlates to a layer thickness of less than 5mm.

Sample taking and manufacturing is difficult. Moisture variations inside the product of more than 2% had been observed.

Positioning the cardboard box between the antennas sometimes gives stupid measuring results, especially when the wall of the box can affect the measurement.

Task list

Field test on tobacco bales in Greece. Bale size should be at least 40 to 50 kg.

Questions: How does the instrument work, when the bale moves?

Which is the most usefull standardisation?

Positioning of the antennas?

Test with waveguides to find out the temperature dependence of the measurement.

Question: Is it necessary and possible to compensate the product temperature?

Test on other products, for instance sand on convayor belt.

Thank you



for your attention!



Harrer & Kassen GmbH
Geschäftsführer Dr. Dipl. Ing. Horst Harrer
Am Heschen 4-6
75328 Langenbrand
Germany
Tel.: +49 (0)7084/9248-0
Fax: +49 (0)7084/924829